



भारतीय खेल
प्राधिकरण

sports
authority of
India

No.: 2(5)/SAI/ES/SSE/2017-18/IFB-145

Date: 06.02.2018


Corrigendum - 02

Sub: Invitation for Bid No. 2(5)/SAI/ES/SSE/2017-18/IFB-145 for Procurement of Medical/Physiotherapy Equipment for NHA, MDC National Stadium, New Delhi.

The following amendments in the specification published along with Bid Document of Section – VI "Technical Specifications" are hereby authorised in the Tender and Bid Document of the subject IFB – 145:-

Item No.	Description of Item/Article	Existing	Amended
3.	Short wave Diathermy	<ul style="list-style-type: none">• Should be Microprocessor Controlled solid state technology.• Should have continuous and Pulsed output.• Should be solid state only.• Should be digital display Without any nobe System.• Should have an output of up to 250W in continuous mode.• Should have an output of up to 500W in Pulsed mode.• Should have a working Frequency of 27.1MHz.• Should have a pulse width from 20 to 400 Micro sec (pulse mode).• Should have inductive and capacitive applicator.• Should have a selectable Treatment time from 1 To 30 minutes.• Should be supplied with All standard accessories Including disc electrodes rubber electrodes, mono – polar inductive electrodes, high frequency cables etc.• Should work with input 200 to 240 Vac 50Hz Supply.	<ul style="list-style-type: none">• Should be Microprocessor Controlled solid state technology.• Should have continuous and Pulsed output.• Should be solid state only.• Should be digital display showing percentage of coupling.• Should have an output of up to 250W in continuous mode.• Should have an output of up to 500W in Pulsed mode.• Should have a working Frequency of 27.1MHz.• Should have a pulse width from 20 to 400 Micro sec (pulse mode).• Should have inductive applicator.• Should have a selectable Treatment time from 1 To 30 minutes.• Should be supplied with All standard accessories Including mono – polar inductive electrodes, high frequency cables etc.• Should work with input 200 to 240 Vac 50Hz Supply.• European CE /US FDA approved.

2. All other entries in the subject IFB will remain unaltered.


(Sanjeev Sharma)
Director (ES)

For and on behalf of DG, SAI

भारतीय खेल प्राधिकरण (युवा कार्यक्रम एवं खेल मंत्रालय)
जवाहरलाल नेहरू स्टेडियम परिसर, पूर्वी द्वार, लोधी रोड, नई दिल्ली-110003
SPORTS AUTHORITY OF INDIA (Ministry of Youth Affairs & Sports)
Jawaharlal Nehru Stadium Complex, East Gate, Lodhi Road, New Delhi-110003
Website : www.sportsauthorityofindia