

**SAI**  
**SPORTS AUTHORITY OF INDIA**  
National Center for Sports Science and Research,  
IGSC, New Delhi – 110002

23.08.2023  
Corrigendum 03

Subject- Amendment to RFP for procuring Cryotherapy Chamber in NCSSR, SAI.


Tender Id - 71-01001(03)/22/2023-NCSSR Division

1. The specification for the procurement stands modified as per the following details.

Technical Specifications as per Tender Document	Specifications after amendments
Should have heat exchange technology so that users never come into contact with cryogenic gases.	Should have heat exchange technology so that users never come into contact with cryogenic gases.
Should have adjustable window enables user comfort	Should have adjustable window enables user comfort
Should have ability to offer Whole Body Cryotherapy treatment that exposes the entire body to extreme temperatures as low as - 140°C in order to promote health & wellness, performance and recovery.	<b>Should have ability to offer Whole Body Cryotherapy treatment that exposes the entire body to extreme temperatures as low as from - 120°C to -140°C in order to promote health &amp; wellness, performance and recovery.</b>
Should have screen that displays the amount of time remaining for the user inside so they can monitor how much time they've already accomplished	Should have screen that displays the amount of treatment time remaining for the user inside so they can monitor how much time they've already accomplished
Should have a convenient low doorstep for easy entry and exit.	Should have a convenient low doorstep for easy entry and exit.
Should have built in oxygen and temperature sensors add an additional level of precaution guaranteeing the safety.	Should have built in oxygen and temperature sensors add an additional level of precaution guaranteeing the safety.
The device should have equipped with a high-quality sound system to play users favourite tracks via Bluetooth and enable audio communication between operator and user	The device should have equipped with a high-quality sound system to play users favourite tracks via Bluetooth and enable audio communication between operator and user
Dimension of Cryo Chamber should be approximately 2500mm(H) x 2000 mm (W) x 1000 mm(D)	<b>Dimension of Cryo Chamber should be approximately 2500mm(H) x 2000 mm (W) x 1000 mm(D) with 20% variation.</b>

Must be supplied cryogenic gas for approximately 50 sessions	Must be supplied cryogenic gas for approximately 50 sessions
Should have European or CE or USFDA safety certificates.	Should have European or CE or USFDA safety certificates.

2. Other Terms and Conditions remained unchanged.

  
Aakash Pundir  
Asstt. Director (NCSSR)

To,

1. All Concerned.

Copy to,

1. Office File