

**NO. 71-01001(03)/23/2023-NCSSR Division for Procurement of TECAR Therapy System
for various NCOEs of SAI the bidders has submitted their queries and the below
are the details**

| Equipment - Tecar Therapy System | | | | | Company Name- |
|---|--|--|--|------------------|----------------------|
| Bio Rehab Technology | | | | | |
| Point No. | Specifications | Representatation | Justification | Amendment | Remarks |
| 2 | Should have 4 functional frequencies range form 300 KHz to 1000 Khz, and upper frequency of 1000 Khz must be available | Frequency Should be Minimum 100W to 400 W | frequency is measured in Hz unit not in watt and Having more frequencies gives the operator the chance to be more specific on the area to treat. Lower frequencies (up to 400 KHz) are efficient for in-depth treatments. Higher frequencies (up to 1000 KHz) are instead normally employed for more superficial | No Amendment | Rejected |

| | | | | | |
|----|--|---|---|--------------|------------------------|
| 4 | Should also Measurement and Indication of the energy have transferred to the tissues | All indication must show on unit screen no need in applicator | Not Applicable | No Amendment | Rejected |
| | Additional Point | Global Installation Report Acceptable | Already Mentioned in tender documents (page 29, SECTION - III -(A)- QUALIFICATION CRITERIA ,Sr. No. 1) Supply and installation outside India will also be considered for evaluation. | No Amendment | As per Tender document |
| 17 | Should Have medical device compliant with EU Regulation 2017/745 or US FDA | Please mention if US FDA/European CE /BIS/CDSO registration as per the new government New Drug Policy MD-14 Required Drug License Policy need MD-41 or 42 for sale any medical equipment | We have already asked US FDA or EU(CE) certificates, if companies have additional quality certificates they can produce it during tender participation | No Amendment | Rejected |
| | | | | | |

| Equipment - Tecar Therapy System | | | | Company | |
|----------------------------------|--|---|--|--------------|----------|
| Name- Cnergy Incorporation | | | | | |
| Point No. | Specifications | Representation | Justification | Amendment | Remarks |
| 6 | Should have automatic modes with minimum 4 output sequences with help of adhesive / Plate electrodes | Should have automatic modes with minimum 3 or more output sequences with help of adhesive / Plate electrodes | All 4 sequences are required and important features for treatment of multiple musculo skeletal conditions in more effective way | No Amendment | Rejected |